



# Target and Vacuum System Status

- Scattering Chamber, supports, cryogenics, sensors, collimator and prototype target have been assembled and installed.
- System has been pumped into the low  $10\text{e-}8$  torr scale.
- Manual valves for roughing system shipped.
- Investigating replacement cost and availability of roughing pump.
- Investigating cost and availability of improved cryo connector for target.
- Getting quotes for parts for symmetric Moller vacuum system.